



Ref. Certif. No.

CH-11560-M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Microbial compressed gas sampler

Name and address of the applicant

MBV AG
Industriestrasse 9, 8712 Stäfa
Switzerland

Name and address of the manufacturer

MBV AG
Industriestrasse 9, 8712 Stäfa
Switzerland

Name and address of the factory

Femron AG
Stadtweg 24, 8245 Feuerthalen
Switzerland

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

MAS-100 Venus: 100-240 VAC; 50-60 Hz 4.5 A max.
MAS-100 Atmos® : 5 VDC, 2.1 A
Internal battery: 7.2 VDC, 10 Ah

Trademark / Brand (if any)

MBV

Customer Test Facility (CTF) Stage used

./.

Model / Type Ref.

MAS-100 Venus
MAS-100 Atmos®

Additional information (if necessary may also be reported on page 2)

National Differences specified in the CB Test Report.
This CBTC replaces CH-11560 (2022-03-16) due to a change of AC/DC adaptor to type SYS1561

A sample of the product was tested and found to be in conformity with

IEC 61326-1:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

20CH-01101.E13, .E14, EMCKP5169A

This CB Test Certificate is issued by the National Certification Body



Eurofins Electric & Electronic Product Testing AG
Luppenstrasse 3
8320 Fehraltorf
SWITZERLAND

Date: 2022-07-20

Signature: Martin Plüss